

REVOCATION OF CERTIFICATES OF APPROVAL ISSUED TO JIANGSU TEYIN NONWOVEN FABRICS CO., LTD. (JTN)

The National Institute for Occupational Safety and Health (NIOSH) has revoked the following certificates of approval issued to Jiangsu Teyin Nonwoven Fabrics Co., Ltd. (JTN):

TC-84A-4244	TC-84A-4245
TC-84A-4380	TC-84A-4381
TC-84A-4382	TC-84A-4384
TC-84A-4385	TC-84A-4386
TC-84A-4389	TC-84A-4406
TC-84A-4407	TC-84A-4408
TC-84A-4490	TC-84A-4501

As of June 22, 2009, no JTN respirator model can be listed as NIOSH approved.

Revocation also means that respirators bearing NIOSH approval numbers TC-84A-4244, TC-84A-4245, TC-84A-4380, TC-84A-4381, TC-84A-4382, TC-84A-4384, TC-84A-4385, TC-84A-4386, TC-84A-4389, TC-84A-4406, TC-84A-4407, TC-84A-4408, TC-84A-4490 and TC-84A-4501 can no longer be manufactured, assembled, sold, or distributed.

NIOSH has revoked these approvals in accordance with the authority granted in Title 42, Code of Federal Regulations, Part 84.43(c) (42 CFR 84) which states "The Institute reserves the right to revoke, for cause, any certificate of approval where it is found that the applicant's quality control test methods, equipment, or records do not ensure effective quality control over the respirator for which approval was issued" and the authority granted in 42 CFR 84.34 which states "The Institute reserves the right to revoke, for cause, any certificate of approval issued pursuant to the provisions of this part. Such causes include, but are not limited to, misuse of approval labels and markings, misleading advertising, and failure to maintain or cause to be maintained the quality control requirements of the certificate of approval.

National Personal Protective Technology Laboratory
 National Institute for Occupational Safety and Health

NIOSH
 PO BOX 18070
 626 COCHRANS MILL ROAD
 PITTSBURGH PA 15236-0070

Telephone 412-386-4000

Fax 412-386-4051

E-mail npptl@cdc.gov

On the Web <http://www.cdc.gov/niosh/npptl/>



NYS Department of State

OFFICE OF
**FIRE PREVENTION
 & CONTROL**

phone 518.474.6746

fax 518.474.3240

email fire@dos.state.ny.us