



## NIOSH HAS REVOKED JTN APPROVAL NUMBERS TC-84A-4404 AND TC-84A-4405

The National Institute for Occupational Safety and Health (NIOSH) wishes to inform respirator users that the NIOSH certificate of approvals TC-84A-4404 and TC-84A-4405 for JTN models MVP-TR1 and MVP-TR2 N95 Filtering Facepieces have been revoked as of December 4, 2008.

These models are being sold by various distributors as NIOSH approved Silver Shield™ N95 Antimicrobial Masks. Complete details concerning the revocation of TC-84A-4404 and TC-84A-4405 can be found at

<http://www.cdc.gov/niosh/npptl/usernotices/default.html>



NYS Department of State

OFFICE OF  
**FIRE PREVENTION  
& CONTROL**

**phone** 518.474.6746  
**fax** 518.474.3240  
**email** fire@dos.state.ny.us